

645—43.9(151) Gonad shielding. Gonad shielding of not less than 0.25 millimeter lead equivalent shall be used for chiropractic patients who have not passed the reproductive age during radiographic procedures in which the gonads are in the useful beam, except for cases in which this would interfere with the diagnostic procedures.