

641—42.11(136C) Information and maintenance record and associated information. The registrant shall maintain all the following information in a separate file or package for each therapeutic radiation machine for inspection by the department:

42.11(1) Report of acceptance testing.

42.11(2) Records of all surveys, calibrations, and periodic quality assurance checks of the therapeutic radiation machine required by rule 641—37.12(136C), as well as the name(s) of person(s) who performed such activities.

42.11(3) Records of maintenance or modifications, or both, performed on the therapeutic radiation machine after July 9, 1997, as well as the name(s) of person(s) who performed such services.

42.11(4) Signature of person authorizing the return of therapeutic radiation machine to clinical use after service, repair, or upgrade.

42.11(5) Records of training specified in rule 641—42.5(136C).

[ARC 0182D, IAB 4/1/26, effective 7/1/26]