

641—42.17(136C) Shielding and safety design requirements.

42.17(1) Each therapeutic radiation machine subject to shall be provided with such primary or secondary barriers as are necessary to ensure compliance with rule 641—37.11(136C).

42.17(2) Facility design information for all new installations of a therapeutic radiation machine or installations of a therapeutic radiation machine of higher energy into a room not previously approved for that energy shall be submitted for department approval prior to actual installation of the therapeutic radiation machine. The minimum facility design information that must be submitted is contained in Appendix B of this chapter.

[ARC 0182D, IAB 4/1/26, effective 7/1/26]