

641—37.6(136C) Prohibited uses. A hand-held fluoroscopic screen cannot be used with X-ray equipment unless it has been accepted for certification by the U.S. Food and Drug Administration, Center for Devices and Radiological Health. A shoe-fitting fluoroscopic device cannot be used. Radiation from radiation-emitting machines or radioactive materials cannot be used on humans for nonmedical purposes except as approved by the department for security-related purposes.

[ARC 0387D, IAB 6/24/26, effective 7/29/26]