

645—63.15(157) Sterilizing instruments. Before use upon a patron in schools and salons, cuticle nippers, tweezers and comedone extractors shall first be cleaned with detergent and water and then sterilized by one of the following methods:

1. Steam sterilizer, registered and listed with the FDA and used according to the manufacturer's instructions;
2. Dry heat sterilizer, registered and listed with the FDA and used according to the manufacturer's instructions; or
3. Sterilization equipment, calibrated to ensure that it reaches the temperature required by manufacturer's instructions.