

21—85.13(215) Master weights. Master scale test weights used for checking scales after being overhauled must be sealed as to their accuracy once every two years. Said weights after being sealed are to be used only as master test weights.

This rule is intended to implement Iowa Code section 215.17.
[ARC 7129C, IAB 12/13/23, effective 1/17/24]