

**657—16.6(155A) Minimum equipment requirements.** Each nuclear pharmacy shall maintain the following equipment for use in the provision of radiopharmaceutical services:

1. Appropriate primary engineering control device to comply with rule 657—16.8(155A);
2. Dose calibrator;
3. Single-channel scintillation counter;
4. Microscope;
5. Incubator, or access to one;
6. Radiation survey meter;
7. Other equipment necessary for the radiopharmaceutical services provided as required by the board.

A pharmacy may request waiver or variance from a provision of this rule pursuant to the procedures and requirements of 657—Chapter 34.

[ARC 3525C, IAB 12/20/17, effective 1/24/18]