

21—85.13(215) Master weights. Master scale test weights used for checking scales after being overhauled must be sealed by the department of agriculture and land stewardship, division of weights and measures, as to their accuracy once each year. Said weights after being sealed are to be used only as master test weights.

This rule is intended to implement Iowa Code section 215.17.